
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001556
11:19 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 2:54
136.0 C 220kPa 2:54
136.4 C 226kPa 3:24
136.4 C 225kPa 3:54
136.3 C 224kPa 4:24
136.4 C 225kPa 4:54
136.4 C 225kPa 5:24
136.0 C 222kPa 5:54
136.4 C 224kPa 6:24
136.4 C 224kPa 6:24
VENTING 6:24
AIR DRYING 7:15
CYCLE COMPLETE 37:15

Digital Signature #
0F248AFBF393C626
