
Midmark M11 - v1.0.5
Total Cycles: 4428

Name: RITTER2
Sterilizer ID: V2313228
03/14/2025 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 71.7 F 0.2
4:00 75.2 F 0.4
6:00 94.0 F 2.0
8:00 142.2 F 5.5
10:00 171.9 F 9.1
12:00 190.9 F 13.1
14:00 205.3 F 4.9
16:00 217.4 F 8.4
18:00 228.5 F 12.4
20:00 230.3 F 8.2
22:00 241.9 F 13.3
24:00 240.8 F 11.3
26:00 252.8 F 17.4
28:00 263.5 F 23.7
30:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.5
2:00 272.5 F 29.9
3:00 271.9 F 29.1
4:00 272.9 F 29.8

Min 271.6 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 181.7 F 0.0
10:00 205.0 F 0.0
15:00 203.4 F 0.0
20:00 198.2 F 0.0
25:00 195.8 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:11
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

03/14/2025 09:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____