
Midmark M11 - v1.0.5
Total Cycles: 1106

Name: RITTER2
Sterilizer ID: V2313228
05/18/2022 07:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 78.1 F 0.0
4:00 111.7 F 0.0
6:00 114.3 F 0.0
8:00 144.7 F 0.0
10:00 185.8 F 0.0
12:00 204.0 F 1.8
14:00 215.2 F 5.0
16:00 224.7 F 8.3
18:00 234.7 F 12.1
20:00 233.0 F 8.5
22:00 243.2 F 12.5
24:00 242.9 F 12.2
26:00 241.9 F 11.3
28:00 254.4 F 17.8
30:00 265.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.4 F 29.0
2:00 272.3 F 29.4
3:00 271.8 F 28.6
4:00 272.7 F 29.4

Min 271.6 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 215.9 F 0.0
10:00 223.5 F 0.0
15:00 224.3 F 0.0
20:00 220.3 F 0.0
25:00 219.7 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:45
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

05/18/2022 08:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____