

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008435  
8:16 24/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:31  
PRESSURIZING 1:36  
STERILIZING 2:41  
136.0 C 220kPa 2:41  
135.7 C 218kPa 3:11  
136.4 C 220kPa 3:41  
136.3 C 219kPa 4:11  
136.4 C 221kPa 4:41  
136.2 C 222kPa 5:11  
136.1 C 220kPa 5:41  
136.2 C 219kPa 6:11  
136.2 C 219kPa 6:11  
VENTING 6:12  
AIR DRYING 7:01  
CYCLE COMPLETE 25:04

-----  
Digital Signature #  
388B36FB27DB5FA3  
-----