
Midmark M11 - v1.0.0
Total Cycles: 3257

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.5 F 0.0
2:00 161.0 F 0.0
4:00 182.2 F 2.9
6:00 201.9 F 7.9
8:00 217.1 F 13.3
10:00 229.5 F 8.2
12:00 241.8 F 13.6
14:00 240.4 F 11.3
16:00 238.9 F 9.8
18:00 252.7 F 16.7
20:00 264.5 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.6 F 29.2
2:00 271.8 F 28.4
3:00 272.7 F 29.2
4:00 271.9 F 28.5

Min 271.6 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 211.8 F 0.0
10:00 224.5 F 0.0
15:00 229.9 F 0.0
20:00 233.4 F 0.0
25:00 227.7 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

05/18/2023 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____