

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4294

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2025 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.4 F 0.0  
2:00 102.9 F 0.8  
4:00 126.0 F 2.9  
6:00 150.0 F 5.7  
8:00 172.9 F 9.0  
10:00 192.9 F 12.3  
12:00 206.6 F 4.3  
14:00 218.3 F 7.7  
16:00 229.1 F 11.6  
18:00 230.8 F 7.6  
20:00 240.9 F 12.5  
22:00 239.9 F 10.7  
24:00 251.2 F 16.4  
26:00 261.3 F 22.3  
28:00 271.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.8  
2:00 272.2 F 29.2  
3:00 271.7 F 29.4  
4:00 272.7 F 29.8

Min 271.5 F 28.9  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 180.3 F 0.0  
10:00 195.9 F 0.0  
15:00 195.6 F 0.0  
20:00 191.4 F 0.0  
25:00 186.5 F 0.0  
30:00 186.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 28:46  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:07:39  
CYCLE COMPLETE

01/24/2025 03:26 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_