
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007492
12:31 04/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:27
STERILIZING 4:22
136.0 C 220kPa 4:22
136.4 C 225kPa 4:53
136.0 C 222kPa 5:23
136.4 C 225kPa 5:53
136.4 C 222kPa 6:23
136.4 C 224kPa 6:53
136.4 C 222kPa 7:23
136.4 C 222kPa 7:53
136.4 C 222kPa 7:53
VENTING 7:53
AIR DRYING 8:48
CYCLE COMPLETE 12:52

Digital Signature #
6FCCB39BC2CF5FB2
