
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002219
11:22 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:27
STERILIZING 3:35
136.4 C 226kPa 3:35
136.3 C 224kPa 4:05
136.3 C 226kPa 4:35
136.4 C 225kPa 5:05
136.3 C 224kPa 5:35
136.2 C 224kPa 6:05
136.2 C 225kPa 6:35
136.4 C 226kPa 7:05
136.4 C 226kPa 7:05
VENTING 7:05
AIR DRYING 8:01
CYCLE COMPLETE 8:21

Digital Signature #
C64176E36CFE3F40
