
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002030
10:34 14/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:06
STERILIZING 4:01
136.0 C 220kPa 4:01
136.0 C 221kPa 4:31
136.3 C 221kPa 5:01
136.4 C 222kPa 5:31
136.1 C 220kPa 6:01
136.1 C 222kPa 6:31
136.4 C 223kPa 7:01
136.4 C 224kPa 7:31
136.4 C 224kPa 7:31
VENTING 7:32
AIR DRYING 8:24
CYCLE COMPLETE 8:32

Digital Signature #
7C990DC7262A90F1
