
Midmark M11 - v1.0.0
Total Cycles: 2167

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 178.8 F 2.0
4:00 186.5 F 6.1
6:00 209.3 F 10.7
8:00 221.6 F 5.1
10:00 233.9 F 10.6
12:00 233.7 F 7.1
14:00 244.0 F 13.5
16:00 242.7 F 11.8
18:00 255.1 F 18.2
20:00 265.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.2 F 29.3
2:00 271.7 F 28.7
3:00 271.7 F 28.9
4:00 272.8 F 29.8

Min 271.0 F 28.4
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 220.1 F 0.0
10:00 232.8 F 0.0
15:00 241.5 F 0.0
20:00 230.2 F 0.0
25:00 221.9 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

07/25/2022 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____