
Midmark M11 - v1.0.5
Total Cycles: 3343

Name: RITTER2
Sterilizer ID: V2313228
02/01/2024 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 78.4 F 0.0
4:00 134.2 F 0.1
6:00 105.9 F 1.3
8:00 141.5 F 4.1
10:00 165.8 F 7.0
12:00 198.5 F 10.1
14:00 206.5 F 7.8
16:00 217.0 F 6.8
18:00 228.5 F 10.7
20:00 228.5 F 7.5
22:00 238.8 F 11.4
24:00 236.6 F 9.4
26:00 249.6 F 15.5
28:00 260.4 F 21.7
30:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.7
2:00 271.4 F 28.8
3:00 272.7 F 29.8
4:00 272.2 F 29.1

Min 271.2 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 231.5 F 0.0
10:00 216.2 F 0.0
15:00 186.3 F 0.0
20:00 182.7 F 0.0
25:00 181.7 F 0.0
30:00 180.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

02/01/2024 10:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____