
Midmark M11 - v1.0.0
Total Cycles: 5375

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2025 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.6 F 0.0
2:00 96.7 F 0.8
4:00 109.7 F 2.7
6:00 145.4 F 5.4
8:00 167.7 F 8.2
10:00 187.5 F 11.5
12:00 206.9 F 3.7
14:00 216.1 F 6.8
16:00 226.7 F 11.0
18:00 228.3 F 6.6
20:00 238.5 F 11.3
22:00 234.6 F 8.6
24:00 247.3 F 14.5
26:00 257.9 F 20.3
28:00 267.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.8
2:00 271.8 F 29.1
3:00 272.4 F 29.9
4:00 272.4 F 29.4

Min 271.5 F 29.1
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 197.2 F 0.0
10:00 191.6 F 0.0
15:00 182.8 F 0.0
20:00 183.7 F 0.0
25:00 199.8 F 0.0
30:00 210.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:25
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

02/27/2025 04:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____