
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001712
7:39 09/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:59
STERILIZING 3:51
136.0 C 220kPa 3:51
136.2 C 222kPa 4:21
136.1 C 219kPa 4:51
136.4 C 222kPa 5:21
136.4 C 222kPa 5:51
136.2 C 218kPa 6:21
136.3 C 222kPa 6:51
136.1 C 218kPa 7:21
136.1 C 218kPa 7:21
VENTING 7:21
AIR DRYING 8:10
CYCLE COMPLETE 8:12

Digital Signature #
758627B30245FCF7
