

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002551  
14:18 28/02/2025  
AIR DRYING ONLY  
60 min. MAXIMUM  
CYCLE START 0:00  
AIR DRYING 0:14  
CYCLE COMPLETE 5:45  
-----

Digital Signature #  
0975A5D5F5203B8F  
-----