

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4042

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2023 02:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.8 F 0.0  
2:00 118.5 F 1.4  
4:00 146.1 F 4.2  
6:00 173.2 F 7.6  
8:00 194.7 F 11.7  
10:00 214.0 F 3.9  
12:00 227.7 F 7.8  
14:00 233.7 F 12.2  
16:00 234.6 F 8.3  
18:00 245.8 F 13.5  
20:00 243.7 F 11.8  
22:00 255.6 F 17.8  
24:00 266.2 F 24.3  
26:00 273.0 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.7  
2:00 271.9 F 28.7  
3:00 272.3 F 28.6  
4:00 271.9 F 28.8

Min 271.4 F 28.2  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 226.8 F 0.0  
10:00 208.4 F 0.0  
15:00 210.6 F 0.0  
20:00 199.8 F 0.0  
25:00 200.8 F 0.0  
30:00 196.4 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 26:24  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:05:06  
CYCLE COMPLETE

11/20/2023 03:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_