
Midmark M11 - v1.0.5
Total Cycles: 173

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.9 F 0.0
2:00 148.5 F 2.5
4:00 169.1 F 6.2
6:00 192.6 F 10.9
8:00 214.0 F 4.8
10:00 223.7 F 10.0
12:00 233.2 F 8.7
14:00 241.4 F 12.9
16:00 242.3 F 11.7
18:00 254.6 F 18.2
20:00 265.1 F 24.9
22:00 271.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 30.0
2:00 271.6 F 29.3
3:00 272.3 F 30.2
4:00 272.5 F 29.7

Min 271.4 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 234.1 F 0.0
10:00 223.5 F 0.0
15:00 187.9 F 0.0
20:00 181.7 F 0.0
25:00 178.4 F 0.0
30:00 176.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

07/07/2021 11:52 AM
ACCEPTED: X
REJECTED:
Operator: U01