
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001937
8:27 09/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:19
STERILIZING 3:24
136.4 C 225kPa 3:24
136.3 C 224kPa 3:54
136.4 C 224kPa 4:24
136.2 C 224kPa 4:54
136.0 C 221kPa 5:24
136.4 C 225kPa 5:54
136.4 C 223kPa 6:24
136.4 C 222kPa 6:54
136.4 C 222kPa 6:54
VENTING 6:54
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
24F7E41DA3B8A1C0
