
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000616
11:27 30/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:40
STERILIZING 3:53
136.0 C 228kPa 3:53
135.9 C 227kPa 4:23
136.3 C 231kPa 4:53
136.4 C 232kPa 5:23
136.5 C 233kPa 5:53
136.3 C 230kPa 6:23
136.3 C 232kPa 6:53
136.3 C 229kPa 7:23
136.3 C 229kPa 7:23
VENTING 7:23
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
23F201426FFEBD9F
