
Midmark M11 - v1.0.0
Total Cycles: 1147

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 222.4 F 0.0
2:00 211.3 F 3.3
4:00 212.1 F 8.6
6:00 229.0 F 9.9
8:00 237.2 F 10.5
10:00 232.7 F 7.6
12:00 241.3 F 10.9
14:00 247.0 F 13.8
16:00 259.6 F 20.7
18:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.6
2:00 272.5 F 29.1
3:00 271.5 F 28.4
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 229.0 F 0.0
10:00 211.0 F 0.0
15:00 208.9 F 0.0
20:00 221.9 F 0.0
25:00 203.9 F 0.0
30:00 166.5 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

09/08/2021 10:13 AM
ACCEPTED: X
REJECTED:
Operator: U01