
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001560
14:55 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:02
PRESSURIZING 6:42
STERILIZING 8:07
136.0 C 220kPa 8:07
136.3 C 221kPa 8:37
136.5 C 223kPa 9:07
136.4 C 222kPa 9:37
136.0 C 222kPa 10:07
136.4 C 219kPa 10:37
136.4 C 222kPa 11:07
136.4 C 223kPa 11:37
VENTING 11:37
AIR DRYING 12:36
CYCLE COMPLETE 22:39

Digital Signature #
E277F211608AE6BA
