
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002197
13:41 09/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:50
STERILIZING 4:50
136.0 C 220kPa 4:50
136.4 C 224kPa 5:20
136.1 C 220kPa 5:50
136.2 C 222kPa 6:20
136.1 C 222kPa 6:50
136.3 C 222kPa 7:20
136.4 C 224kPa 7:50
136.3 C 224kPa 8:20
136.3 C 224kPa 8:20
VENTING 8:20
AIR DRYING 9:20
CYCLE COMPLETE 9:22

Digital Signature #
46CBEDC49B7364E9
