
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000625
9:38 02/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:57
PRESSURIZING 3:59
STERILIZING 5:17
136.0 C 228kPa 5:17
136.4 C 231kPa 5:47
136.3 C 230kPa 6:17
136.3 C 230kPa 6:47
136.4 C 232kPa 7:17
136.4 C 232kPa 7:47
136.4 C 232kPa 8:17
136.4 C 232kPa 8:47
136.4 C 232kPa 8:47
VENTING 8:47
AIR DRYING 9:46
CYCLE COMPLETE 24:46

Digital Signature #
6F27E6BAA0E49877
