
Midmark M11 - v1.0.5
Total Cycles: 140

Name: RITTER2
Sterilizer ID: V2313228
06/24/2021 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.3 F 0.0
2:00 119.6 F 1.4
4:00 136.0 F 4.3
6:00 165.8 F 8.6
8:00 196.2 F 10.9
10:00 210.8 F 7.8
12:00 226.4 F 13.2
14:00 233.1 F 10.6
16:00 236.2 F 9.1
18:00 249.8 F 15.8
20:00 260.1 F 22.6
22:00 271.0 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.2
1:00 272.3 F 30.1
2:00 273.1 F 30.3
3:00 273.3 F 30.6
4:00 272.6 F 30.8

Min 271.9 F 30.1
Max 273.6 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 233.4 F 0.0
10:00 226.6 F 0.0
15:00 227.7 F 0.0
20:00 218.5 F 0.0
25:00 191.3 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:33
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

06/24/2021 03:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____