
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003338
10:05 30/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:26
136.4 C 221kPa 3:26
136.4 C 225kPa 3:56
136.1 C 223kPa 4:26
136.3 C 222kPa 4:56
136.4 C 225kPa 5:26
136.4 C 224kPa 5:56
136.4 C 223kPa 6:26
136.4 C 222kPa 6:56
136.4 C 222kPa 6:56
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 17:16

Digital Signature #
3D7FE638313E8B73
