
Midmark M11 - v1.0.5
Total Cycles: 1656

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 169.8 F 2.5
4:00 170.8 F 6.5
6:00 217.4 F 12.2
8:00 221.2 F 6.6
10:00 231.8 F 11.6
12:00 232.7 F 8.6
14:00 239.8 F 11.0
16:00 245.6 F 13.4
18:00 257.2 F 19.8
20:00 268.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.2 F 29.3
2:00 272.5 F 29.6
3:00 272.3 F 29.9
4:00 271.6 F 29.3

Min 271.4 F 29.2
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 233.5 F 0.0
10:00 228.5 F 0.0
15:00 228.9 F 0.0
20:00 224.4 F 0.0
25:00 231.2 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

10/25/2022 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____