
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000368
8:35 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:38
STERILIZING 3:44
136.3 C 220kPa 3:44
136.2 C 230kPa 4:14
136.2 C 232kPa 4:44
136.4 C 233kPa 5:14
136.4 C 233kPa 5:44
136.4 C 233kPa 6:14
136.2 C 229kPa 6:44
136.3 C 233kPa 7:14
136.3 C 233kPa 7:14
VENTING 7:14
AIR DRYING 8:05
CYCLE COMPLETE 8:14

Digital Signature #
35816A29F924AB6E
