
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001470
10:16 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:07
STERILIZING 4:13
136.5 C 225kPa 4:13
136.3 C 225kPa 4:43
136.4 C 225kPa 5:13
136.1 C 224kPa 5:43
136.4 C 226kPa 6:13
136.3 C 224kPa 6:43
136.4 C 226kPa 7:13
136.3 C 225kPa 7:43
136.3 C 225kPa 7:43
VENTING 7:43
AIR DRYING 8:37
CYCLE COMPLETE 11:19

Digital Signature #
8E7A84EEB930CC68
