
Midmark M11 - v1.0.0
Total Cycles: 2088

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 173.2 F 2.4
4:00 177.9 F 6.3
6:00 206.3 F 11.1
8:00 221.7 F 6.0
10:00 231.6 F 11.5
12:00 232.7 F 8.0
14:00 241.4 F 11.4
16:00 244.0 F 12.4
18:00 256.7 F 19.0
20:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.7 F 29.6
2:00 271.9 F 28.9
3:00 272.6 F 29.5
4:00 271.8 F 29.1

Min 270.9 F 28.3
Max 273.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 230.6 F 0.0
10:00 231.0 F 0.0
15:00 228.9 F 0.0
20:00 229.5 F 0.0
25:00 230.3 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

07/05/2022 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____