
Midmark M11 - v1.0.0
Total Cycles: 1578

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.8 F 0.0
2:00 86.3 F 0.0
4:00 90.5 F 0.0
6:00 116.3 F 2.0
8:00 146.3 F 5.1
10:00 175.0 F 8.8
12:00 194.2 F 12.7
14:00 209.4 F 4.1
16:00 221.2 F 7.8
18:00 231.0 F 11.8
20:00 231.8 F 7.7
22:00 244.0 F 13.1
24:00 242.3 F 11.3
26:00 254.4 F 17.5
28:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 28.8
2:00 271.6 F 28.2
3:00 272.8 F 28.9
4:00 271.6 F 28.1

Min 271.5 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 231.1 F 0.0
10:00 226.5 F 0.0
15:00 220.9 F 0.0
20:00 218.3 F 0.0
25:00 217.8 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:07
CYCLE COMPLETE

02/01/2022 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____