
Midmark M11 - v1.0.5
Total Cycles: 1327

Name: RITTER2
Sterilizer ID: V2313228
07/21/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 74.3 F 0.0
4:00 78.5 F 0.0
6:00 108.0 F 0.3
8:00 142.3 F 2.8
10:00 170.8 F 5.7
12:00 187.6 F 9.5
14:00 205.0 F 13.3
16:00 217.0 F 6.1
18:00 227.3 F 10.1
20:00 232.0 F 9.5
22:00 237.9 F 11.3
24:00 235.6 F 9.2
26:00 249.1 F 15.5
28:00 260.0 F 21.7
30:00 270.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.2 F 29.5
2:00 271.8 F 29.5
3:00 272.0 F 29.3
4:00 271.9 F 29.6

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 231.5 F 0.0
10:00 200.8 F 0.0
15:00 159.5 F 0.0
20:00 156.3 F 0.0
25:00 145.9 F 0.0
30:00 159.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

07/21/2022 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____