
Midmark M11 - v1.0.0
Total Cycles: 3157

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.3 F 0.0
2:00 84.6 F 0.0
4:00 92.6 F 0.0
6:00 152.4 F 0.0
8:00 148.1 F 0.0
10:00 184.0 F 0.4
12:00 208.6 F 3.9
14:00 228.6 F 6.8
16:00 235.2 F 10.1
18:00 237.7 F 10.6
20:00 238.9 F 9.8
22:00 238.5 F 9.4
24:00 245.6 F 12.7
26:00 242.2 F 11.0
28:00 254.7 F 17.4
30:00 265.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.5 F 28.4
2:00 271.4 F 27.9
3:00 272.6 F 28.5
4:00 271.5 F 27.8

Min 271.4 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 198.6 F 0.0
10:00 216.7 F 0.0
15:00 232.5 F 0.0
20:00 225.8 F 0.0
25:00 226.0 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:41
CYCLE COMPLETE

04/25/2023 08:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____