
Midmark M11 - v1.0.0
Total Cycles: 5120

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2024 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.6 F 0.0
2:00 141.0 F 0.0
4:00 167.6 F 0.3
6:00 192.1 F 0.8
8:00 215.3 F 4.1
10:00 227.1 F 8.0
12:00 238.3 F 12.2
14:00 234.6 F 8.7
16:00 241.1 F 11.5
18:00 243.8 F 12.4
20:00 241.9 F 11.3
22:00 254.2 F 17.6
24:00 265.2 F 24.0
26:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.2 F 28.9
2:00 272.7 F 28.9
3:00 272.1 F 28.5
4:00 271.4 F 28.2

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 192.8 F 0.0
10:00 194.0 F 0.0
15:00 193.7 F 0.0
20:00 190.4 F 0.0
25:00 187.0 F 0.0
30:00 184.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 26:00
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:42
CYCLE COMPLETE

11/21/2024 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____