

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3658

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2023 03:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
2:00 201.3 F 3.9  
4:00 207.3 F 9.2  
6:00 222.3 F 7.3  
8:00 232.7 F 10.3  
10:00 233.4 F 7.8  
12:00 243.0 F 11.7  
14:00 246.1 F 13.3  
16:00 259.1 F 20.3  
18:00 270.6 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 271.6 F 28.4  
2:00 271.7 F 28.3  
3:00 271.8 F 28.3  
4:00 272.0 F 28.5  
  
Min 271.6 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 215.4 F 0.0  
10:00 228.9 F 0.0  
15:00 230.7 F 0.0  
20:00 228.7 F 0.0  
25:00 226.9 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 2:29  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

08/15/2023 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_