
Midmark M11 - v1.0.0
Total Cycles: 3403

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.0 F 0.0
2:00 131.6 F 0.0
4:00 149.6 F 0.9
6:00 164.1 F 3.4
8:00 186.4 F 7.2
10:00 203.5 F 11.5
12:00 216.9 F 4.7
14:00 229.9 F 9.3
16:00 237.5 F 10.8
18:00 240.1 F 11.1
20:00 236.7 F 8.9
22:00 250.4 F 15.5
24:00 261.8 F 21.9
26:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.1
2:00 271.6 F 28.2
3:00 272.2 F 29.1
4:00 272.1 F 28.7

Min 271.2 F 28.1
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 212.1 F 0.0
10:00 223.5 F 0.0
15:00 220.7 F 0.0
20:00 221.1 F 0.0
25:00 220.8 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 26:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

06/19/2023 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____