
Midmark M11 - v1.0.0
Total Cycles: 1175

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.2 F 0.0
2:00 175.8 F 3.2
4:00 195.4 F 8.0
6:00 231.5 F 10.5
8:00 229.6 F 9.9
10:00 235.6 F 7.9
12:00 243.1 F 11.7
14:00 245.4 F 13.1
16:00 258.8 F 20.3
18:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 271.8 F 28.9
2:00 272.4 F 29.2
3:00 272.9 F 29.5
4:00 272.0 F 29.4

Min 271.5 F 28.9
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 243.6 F 0.0
10:00 230.4 F 0.0
15:00 208.9 F 0.0
20:00 193.7 F 0.0
25:00 195.1 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

09/15/2021 10:15 AM
ACCEPTED: X
REJECTED:
Operator: U01