

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001629  
13:50 26/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:33  
STERILIZING 3:25  
136.0 C 220kPa 3:25  
136.4 C 224kPa 3:55  
136.4 C 222kPa 4:25  
136.1 C 223kPa 4:55  
136.4 C 225kPa 5:25  
136.2 C 219kPa 5:55  
136.4 C 224kPa 6:25  
136.4 C 224kPa 6:55  
136.4 C 224kPa 6:55  
VENTING 6:55  
AIR DRYING 7:47  
CYCLE COMPLETE 7:53

-----  
Digital Signature #  
8C779EFDB90F24D1  
-----