
Midmark M11 - v1.0.0
Total Cycles: 2401

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2022 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 177.7 F 2.0
4:00 180.2 F 5.7
6:00 209.2 F 10.5
8:00 224.0 F 5.4
10:00 236.9 F 11.1
12:00 233.1 F 8.0
14:00 242.0 F 11.0
16:00 245.3 F 13.0
18:00 258.3 F 19.8
20:00 269.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.4
3:00 272.4 F 29.1
4:00 271.6 F 28.4

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 232.1 F 0.0
10:00 237.6 F 0.0
15:00 235.5 F 0.0
20:00 223.7 F 0.0
25:00 221.7 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

10/05/2022 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____