
Midmark M11 - v1.0.5
Total Cycles: 3569

Name: RITTER2
Sterilizer ID: V2313228
04/29/2024 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.2 F 0.0
2:00 85.7 F 0.2
4:00 103.5 F 1.7
6:00 138.6 F 4.2
8:00 160.7 F 7.2
10:00 179.9 F 10.5
12:00 204.7 F 6.3
14:00 209.7 F 6.4
16:00 221.4 F 10.4
18:00 228.7 F 8.4
20:00 236.4 F 11.2
22:00 235.3 F 9.1
24:00 248.2 F 15.1
26:00 258.9 F 21.2
28:00 269.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.9 F 30.3
2:00 272.3 F 29.8
3:00 271.2 F 29.4
4:00 272.1 F 30.2

Min 271.2 F 29.3
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 214.2 F 0.0
10:00 203.1 F 0.0
15:00 199.9 F 0.0
20:00 203.1 F 0.0
25:00 202.7 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:11
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:42
CYCLE COMPLETE

04/29/2024 03:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____