

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4124

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.8 F 0.0  
2:00 178.5 F 2.8  
4:00 221.1 F 7.1  
6:00 216.6 F 12.0  
8:00 225.0 F 6.0  
10:00 236.1 F 10.8  
12:00 233.7 F 7.4  
14:00 245.7 F 13.4  
16:00 244.3 F 12.2  
18:00 256.5 F 18.5  
20:00 267.3 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.6  
2:00 272.6 F 28.7  
3:00 272.4 F 28.8  
4:00 272.2 F 28.9

Min 270.9 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 185.5 F 0.0  
10:00 198.8 F 0.0  
15:00 199.0 F 0.0  
20:00 196.3 F 0.0  
25:00 192.6 F 0.0  
30:00 190.4 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 21:20  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:51  
CYCLE COMPLETE

12/15/2023 12:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_