

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3474

Name: RITTER2  
Sterilizer ID: V2313228  
03/25/2024 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.9 F 0.0  
2:00 113.2 F 0.1  
4:00 118.9 F 1.2  
6:00 146.7 F 3.9  
8:00 175.8 F 7.3  
10:00 193.8 F 10.9  
12:00 205.0 F 4.6  
14:00 218.8 F 7.4  
16:00 231.0 F 11.9  
18:00 233.9 F 8.4  
20:00 242.3 F 12.0  
22:00 244.3 F 12.6  
24:00 255.6 F 18.6  
26:00 266.1 F 25.1  
28:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.2  
1:00 272.5 F 29.9  
2:00 272.5 F 29.5  
3:00 271.5 F 29.3  
4:00 272.8 F 29.9

Min 271.4 F 29.1  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.0 F 0.0  
5:00 193.8 F 0.0  
10:00 196.0 F 0.0  
15:00 197.3 F 0.0  
20:00 193.5 F 0.0  
25:00 189.7 F 0.0  
30:00 188.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:23  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:06:53  
CYCLE COMPLETE

03/25/2024 11:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_