
Midmark M11 - v1.0.5
Total Cycles: 416

Name: RITTER2
Sterilizer ID: V2313228
10/05/2021 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 74.0 F 0.0
4:00 76.1 F 0.0
6:00 92.1 F 0.3
8:00 122.8 F 2.7
10:00 150.2 F 5.7
12:00 171.8 F 9.2
14:00 193.1 F 10.9
16:00 206.6 F 7.3
18:00 219.8 F 11.3
20:00 227.3 F 7.4
22:00 240.3 F 12.8
24:00 239.3 F 10.8
26:00 251.0 F 16.7
28:00 261.5 F 22.6
30:00 271.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 272.4 F 30.0
2:00 271.4 F 29.5
3:00 272.2 F 29.6
4:00 271.8 F 30.0

Min 271.2 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 203.9 F 0.0
10:00 201.2 F 0.0
15:00 198.2 F 0.0
20:00 196.4 F 0.0
25:00 196.6 F 0.0
30:00 175.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:34
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

10/05/2021 09:03 AM
ACCEPTED: X
REJECTED:
Operator: U01