
Midmark M11 - v1.0.0
Total Cycles: 1268

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.3 F 0.0
2:00 183.5 F 3.4
4:00 205.7 F 8.5
6:00 229.9 F 9.4
8:00 235.8 F 9.4
10:00 238.0 F 9.1
12:00 245.1 F 12.8
14:00 243.0 F 11.6
16:00 256.4 F 18.5
18:00 268.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.1 F 28.5
2:00 271.9 F 28.7
3:00 271.8 F 28.4
4:00 271.9 F 28.7

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 202.9 F 0.0
10:00 202.7 F 0.0
15:00 204.2 F 0.0
20:00 202.7 F 0.0
25:00 200.6 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

10/14/2021 01:46 PM
ACCEPTED: X
REJECTED:
Operator: U02