
Midmark M11 - v1.0.0
Total Cycles: 1532

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.2 F 0.0
2:00 145.5 F 2.0
4:00 161.2 F 5.6
6:00 185.3 F 10.0
8:00 204.2 F 4.7
10:00 219.2 F 7.7
12:00 232.2 F 12.5
14:00 234.5 F 9.5
16:00 236.9 F 9.2
18:00 246.2 F 13.9
20:00 257.9 F 20.3
22:00 268.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.6
2:00 271.6 F 29.0
3:00 272.6 F 29.7
4:00 271.8 F 29.1

Min 271.5 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 209.6 F 0.0
10:00 218.6 F 0.0
15:00 230.9 F 0.0
20:00 228.7 F 0.0
25:00 225.9 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:10
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

01/17/2022 01:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____