
Midmark M11 - v1.0.0
Total Cycles: 3870

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 189.4 F 3.0
4:00 199.7 F 7.4
6:00 208.9 F 12.5
8:00 220.8 F 6.7
10:00 232.9 F 11.6
12:00 233.4 F 8.5
14:00 241.7 F 11.1
16:00 245.7 F 13.2
18:00 257.7 F 19.6
20:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.1 F 28.7
2:00 272.4 F 29.2
3:00 271.5 F 28.4
4:00 272.5 F 29.0

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 213.3 F 0.0
10:00 220.3 F 0.0
15:00 227.6 F 0.0
20:00 228.8 F 0.0
25:00 231.7 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

10/05/2023 10:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____