

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3250

Name: RITTER2  
Sterilizer ID: V2313228  
12/22/2023 11:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.7 F 0.0  
2:00 159.7 F 2.1  
4:00 175.2 F 5.6  
6:00 196.0 F 10.1  
8:00 224.0 F 6.2  
10:00 221.2 F 9.3  
12:00 232.1 F 9.7  
14:00 241.5 F 11.8  
16:00 239.8 F 10.4  
18:00 253.0 F 16.9  
20:00 264.1 F 23.6  
22:00 270.9 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.6  
2:00 271.5 F 29.2  
3:00 272.8 F 29.6  
4:00 271.3 F 29.0

Min 271.3 F 28.9  
Max 273.2 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
5:00 238.9 F 0.0  
10:00 196.8 F 0.0  
15:00 213.6 F 0.0  
20:00 209.2 F 0.0  
25:00 193.9 F 0.0  
30:00 178.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 22:46  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:19  
CYCLE COMPLETE

12/22/2023 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_