
Midmark M11 - v1.0.5
Total Cycles: 2758

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 164.3 F 0.0
4:00 183.4 F 0.0
6:00 201.9 F 0.6
8:00 214.1 F 4.4
10:00 225.3 F 8.5
12:00 236.3 F 13.0
14:00 237.2 F 10.2
16:00 237.3 F 9.2
18:00 241.7 F 11.0
20:00 248.1 F 14.4
22:00 259.6 F 20.8
24:00 270.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.4 F 28.9
2:00 272.4 F 29.3
3:00 271.7 F 28.5
4:00 272.7 F 29.2

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 223.5 F 0.0
10:00 226.5 F 0.0
15:00 224.2 F 0.0
20:00 218.1 F 0.0
25:00 214.5 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 24:54
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:03:43
CYCLE COMPLETE

08/14/2023 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____