
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001415
9:29 29/03/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:47
136.4 C 220kPa 4:47
136.2 C 223kPa 5:17
136.4 C 222kPa 5:47
136.4 C 224kPa 6:17
136.3 C 223kPa 6:47
136.2 C 223kPa 7:17
136.2 C 222kPa 7:47
136.4 C 224kPa 8:17
136.4 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 9:14

Digital Signature #
D9F7C5C2680DECD6
