
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001345
13:38 11/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:54
136.0 C 221kPa 3:54
136.1 C 223kPa 4:24
136.4 C 223kPa 4:54
136.3 C 222kPa 5:24
136.4 C 225kPa 5:54
136.3 C 225kPa 6:24
136.4 C 226kPa 6:54
136.5 C 226kPa 7:24
136.5 C 226kPa 7:24
VENTING 7:24
AIR DRYING 8:16
CYCLE COMPLETE 8:36

Digital Signature #
011AE12B00685336
