
Midmark M11 - v1.0.0
Total Cycles: 2104

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2022 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.6 F 0.0
2:00 88.1 F 0.0
4:00 102.0 F 0.8
6:00 132.7 F 2.5
8:00 160.2 F 4.8
10:00 182.7 F 7.9
12:00 209.2 F 11.5
14:00 218.3 F 4.5
16:00 226.0 F 8.3
18:00 236.7 F 12.4
20:00 234.2 F 8.7
22:00 242.6 F 11.7
24:00 243.2 F 12.1
26:00 254.8 F 18.0
28:00 265.2 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.2
2:00 272.7 F 29.0
3:00 271.7 F 28.4
4:00 271.5 F 28.6

Min 271.3 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 230.2 F 0.0
10:00 237.3 F 0.0
15:00 218.3 F 0.0
20:00 222.5 F 0.0
25:00 229.7 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:60
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:04
CYCLE COMPLETE

07/08/2022 10:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____