
Midmark M11 - v1.0.5
Total Cycles: 3301

Name: RITTER2
Sterilizer ID: V2313228
01/15/2024 04:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.5 F 0.0
2:00 184.4 F 3.6
4:00 209.1 F 9.7
6:00 224.7 F 6.3
8:00 238.4 F 12.0
10:00 239.1 F 10.5
12:00 238.5 F 9.9
14:00 254.5 F 17.8
16:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.4 F 28.9
2:00 272.7 F 29.4
3:00 271.8 F 28.5
4:00 272.8 F 29.4

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 236.7 F 0.0
10:00 224.2 F 0.0
15:00 224.2 F 0.0
20:00 223.4 F 0.0
25:00 218.4 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 17:12
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:56:45
CYCLE COMPLETE

01/15/2024 05:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____